



## EC Declaration of conformity no.: 190625-001 (issue 1)

Product Name:	Motion Restriction Pillow MRP Vacuum Pillow, Ferno
Germa Article No:	15 000 000 114
Description:	<p>Intended use is to give comfort and support patient head during ambulance transport on top of a stretcher in medical preferred position, using vacuum technology. Fully welded product including a valve and to be evacuated by an evacuation pump.</p> <p>Designed to be used on top of a Germa stretcher mattress and fixed on the framework of the stretcher top.</p> <p>Size: 36cm H x 48cm W x 10cm T Weight: 0,7 Kg</p>
Accessory:	

Identification: All products with serial numbers issued after LOT number: **030704**, date: **2003-07-04** (yyyy-mm-dd).

We the manufacturer hereby declare that this product is in conformity with the requirements in:

- Annex I and Annex VII and Annex IX in Council directive 93/42/EEC. Equipment class: I, non-sterile, no measuring function.
- Swedish Medical Devices Act 1993:584

Signed in Kristianstad, Sweden on behalf of AB Germa by:

Position: Acting Quality Manager  
Name: Björn Holmqvist

Date: 25/08 2019

Sign: 

Postadress	Adress	Telefon	Telefax	Organisationsnummer.	Bankgiro	Postgiro
Box 2127				F-skattebevis		
S-291 02 KRISTIANSTAD SWEDEN	Industrigatan 54-56 KRISTIANSTAD	+46 (0)44-123030	+46 (0)44-103179	556086-9413	481-4125	21 68 22-7
e-mail: info@germa.se						